

**Laboratory Accreditation Programs  
List of Deficiencies**

**LAP Number:** 8863194  
**Accreditation Unit (AU):** Advanta Toxicology, LLC-Laboratory  
**Address:** 10935 County Road 159  
Tyler, TX 75703

**Director:** Tatum, Owatha L., PhD,MBA,HCLD,CC(ABB)  
**AU Inspection Date:** 08/15/2016  
**Inspection Type:** Initial  
**Team Leader:** Weissfeld, Alice S., PhD  
**Number of Deficiencies:** 2

**Laboratory Accreditation Programs  
List of Deficiencies**

**AU ID:** 1736202  
**Accreditation Unit (AU):** Advanta Toxicology, LLC-Laboratory

**SU ID:** 1792333  
**Section Unit (SU):** Clinical Toxicology  
**SU Inspection Date:** 08/15/2016

<b>Checklist</b>	<b>Requirement ID</b>	<b>Phase</b>	<b>Requirement</b>
All Common	COM.01400	II	The proficiency testing attestation statement is signed by the laboratory director or designee and all individuals involved in the testing process.

**Laboratory Accreditation Programs**  
**List of Deficiencies**

**AU ID:** 1736202  
**Accreditation Unit (AU):** Advanta Toxicology, LLC-Laboratory  
**SU ID:** 1788769  
**Section Unit (SU):** Lab General  
**SU Inspection Date:** 08/15/2016

<b>Checklist</b>	<b>Requirement ID</b>	<b>Phase</b>	<b>Requirement</b>
------------------	-----------------------	--------------	--------------------

---

**There were no deficiencies from the inspection of this section.**

**Laboratory Accreditation Programs  
List of Deficiencies**

**AU ID:** 1736202  
**Accreditation Unit (AU):** Advanta Toxicology, LLC-Laboratory

**SU ID:** 1792335  
**Section Unit (SU):** Molecular Pathology  
**SU Inspection Date:** 08/15/2016

<b>Checklist</b>	<b>Requirement ID</b>	<b>Phase</b>	<b>Requirement</b>
Microbiology	MIC.64760	II	There is documentation that the laboratory has performed a verification study prior to reporting patient results.